



QPS Evaluation Services Inc
Testing, Certification and Field Evaluation Body
Accredited in Canada, the USA, and Internationally

| |
|--------|
| File |
| LR3030 |

CERTIFICATE OF COMPLIANCE
(ISO TYPE 3 CERTIFICATION SYSTEM)

| | |
|--------------------------------|---|
| Issued to | Heron Instruments Inc. |
| Address | 447 Moxley Road - Dundas, Ontario L9H 5E2 - Canada |
| Project Number | LR3030-1 |
| Product | Portable Oil Water Interface Meter |
| Model Number | H.OIL followed with 1602-xxxxM or 1602-xxxxF, where xxxx stands for tape length, M for meter (1500 max) and F for feet (4920 max). |
| Electrical Ratings | Powered by one 9V Alkaline battery Duracell MN1604, Procell PC1604 or Energizer 522. |
| Markings | Class I, Division 1, Group CD, T3 Class I, Zone 0, AEx ia IIB T3 Ga Ex ia IIB T3 Ga Ta - Probe: -45°C to +135°C Reel: -45°C to +55°C |
| Applicable Standards | CSA 60079-0:2019 UL 60079-0 Seventh Ed. CSA 60079-11:2014 UL 60079-11 Sixth Ed. |
| Factory/Manufacturing Location | Same as above. |
| Conditions of Safe Use | Use only one of the following 9V Alkaline cells: Duracell MN1604, Procell PC1604 or Energizer 522. Must use ground clamp to reduce the risk of electrostatic charge. |

Statement of Compliance: The product(s)/equipment identified in this Certificate and described in the Certification Report covered under the above referenced project number have been investigated and found to be in compliance with the relevant requirements of the above referenced standard(s). As such, they are eligible to bear the QPS Certification Mark shown below, in accordance with the provisions of QPS's Service Agreement.

IMPORTANT NOTE: In order to maintain the integrity of the QPS Mark(s), certification will be revoked if:

- (1) Compliance to the above-mentioned Standard(s), or those identified in future QPS Standard Update Notice – SUN (QSD 55) is not maintained, or,
- (2) If the product/equipment is modified after certification is granted without prior written consent by QPS



Issued By: Rob Kohuch, Senior Engineer

Signature: *Rob Kohuch*

Date: November 9, 2023



The SCC and IAS Accreditation Symbols are official symbols of the respective accreditation bodies, used under license.

81 Kelfield St., Unit 8, Toronto, ON M9W 5A3 Tel: 416-241-8857; Fax: 416-241-0682

www.qps.ca

